

Cleveland State University

EngagedScholarship@CSU

---

Undergraduate Research Posters 2016

Undergraduate Research Posters

---

2016

## Wrist-Mounted Feedback System: Monitoring Force and Torque in Six Degrees of Freedom

Christopher Schroeck  
*Cleveland State University*

Follow this and additional works at: [https://engagedscholarship.csuohio.edu/u\\_poster\\_2016](https://engagedscholarship.csuohio.edu/u_poster_2016)



Part of the [Engineering Commons](#)

[How does access to this work benefit you? Let us know!](#)

---

### Recommended Citation

Schroeck, Christopher, "Wrist-Mounted Feedback System: Monitoring Force and Torque in Six Degrees of Freedom" (2016). *Undergraduate Research Posters 2016*. 59.

[https://engagedscholarship.csuohio.edu/u\\_poster\\_2016/59](https://engagedscholarship.csuohio.edu/u_poster_2016/59)

This Book is brought to you for free and open access by the Undergraduate Research Posters at EngagedScholarship@CSU. It has been accepted for inclusion in Undergraduate Research Posters 2016 by an authorized administrator of EngagedScholarship@CSU. For more information, please contact [library.es@csuohio.edu](mailto:library.es@csuohio.edu).



# ***Wrist-Mounted Feedback System: Monitoring Force and Torque in Six Degrees of Freedom***

Washkewicz College of Engineering

**Student Researcher:** Christopher Schroeck

**Faculty Advisor:** Eric Schearer

## **Abstract**

This project developed a wrist-mountable, six Degrees of Freedom (DoF) feedback system that monitors force and torque. This system is designed to be used by nontechnical operators to correct movements of a paralyzed arm under the control of a functional electrical stimulation (FES) neuroprosthesis. The resulting corrections are recorded and processed by additional programs so as to improve the accuracy of future movements controlled by the FES system. The feedback system was tested on a robotic arm that is programmed to simulate a paralyzed human's arm. These tests checked the sensor for accuracy, reliability, durability, ease of use, and stability.

After production and testing, the feedback system was found to be durable enough to repeatedly withstand the forces of moving the robotically simulated paralyzed arm. The system was considered easy preform corrections with by testers after having minimal to no prior instruction and can readily record forces and torques input during the correction process, accurate within  $\pm 1\text{N}$ .